Safety instructions VEGABOX 03

Intrinsic safety "i"





Document ID: 47691







Contents

1	Area of applicability	. 3
2	Important specification in the type code	. 3
3	General information	. 4
4	Application area	. 4
5	Specific conditions of use ("X" identification)	. 4
6	Important information for mounting and maintenance	. 5
7	Safe operating mode	. 6
8	Potential equalization/Grounding	. 6
9	Electrostatic charging (ESD)	. 6
10	Instructions for zone 0 applications	. 7
11	Electrical data	. 7
12	Thermal data	. 8

Supplementary documentation:

- Operating Instructions VEGABOX 03
 Certificate of Conformity IECEx TUR 14.0007X (Document ID: 47692)

Editing status: 2021-04-16



1 Area of applicability

These safety instructions apply to the VEGABOX 03 of type series:

- BOX03(*).IC*******
- BOX03(*).IO*******

According to Certificate of Conformity IECEx TUR 14.0007X (certificate number on the type label) and for all instruments with safety instruction 47691.

The classification as well as the respective standards are stated in the Certificate of Conformity:

- IEC 60079-0: 2017
- IEC 60079-11:2011
- IEC 60079-26: 2021

Type of protection marking:

- Ex ia IIC T6 ... T1 Ga
- Ex ia IIC T6 ... T1 Gb

2 Important specification in the type code

BOX03(*).abcdefghi

Pos	sition	Feature	Description
а	Scope	I	Worldwide
b	Approval	С	Ex ia IIC T6 T1 Ga, Gb
b	Approvai	0	Ex ia IIC T6 T1 Ga, Gb + Ship approval (DNV GL, BV, RMROS)
с	Version	А	Pressure compensation
d	Housing	к	Plastic
		А	Aluminium
		V	Stainless steel (precision casting)
		н	Special colour, Aluminium
е	Protection rating	I	IP66/IP67; NEMA 4X
		N	IP66/IP68 (0.2 bar); NEMA 6P
f	Cable entry / Connection	D	M20 x 1.5 / Blind plug
		N	1/2 NPT / Blind plug
		М	M20 x 1.5 / Cable gland PA black (ø5-9 mm), standard
		J	1/2 NPT / Cable gland PA black (ø5-9 mm)
		*	further cable glands, blind plugs, cable leadthroughs, plug connectors, Conduit system
		A	for wall mounting with Aluminium or stainless steel housing
2	Mounting type	С	for carrier rail and wall mounting with plastic housing
g		D	for carrier rail with Aluminium or stainless steel housing
		E	for tube mounting (29 60 mm) incl. mounting material
h	Certificates	х	No
11	Certificates	М	Yes

47691-EN-210607

In the following, all above mentioned versions are called VEGABOX 03. If parts of these safety



instructions refer only to certain versions, then these will be mentioned explicitly with their type code.

3 General information

The VEGABOX 03 is preferably used for field mounting for separated connection of sensor circuits and as breather housing. The VEGABOX 03 of type series VEGABOX 03 with integrated connection terminals are preferably used for pressure compensation of the pressure measuring cell and as terminal box in conjunction with pressure transmitters of Messrs. VEGA in the cable version with capillary cable.

In VEGABOX 03 only terminal blocks as type VEGABOX 03 for connection of intrinsically safe circuits can be installed. A terminal block is preferably used for connection of an inrinsically safe circuit of VEGA pressure transmitters in the version with connection cable with corresponding power supply or signal conditioning instrument.

The VEGABOX 03 is an intrinsically safe electrical instrument for installation in hazardous areas with combustible gases, mist or vapour, requiring instruments of category EPL Ga or EPL Gb or for installation outside of hazardous areas. The VEGABOX 03 is an intrinsically safe instrument for installation in hazardous areas of all combustible materials of explosion group IIA, IIB and IIC.

If the VEGABOX 03 are installed and operated in hazardous areas, the general Ex installation regulations IEC 60079-14 as well as these safety instructions must be observed.

The operating instructions as well as the installation regulations or standards that apply for explosion protection of electrical systems must generally be observed.

The installation of explosion-endangered systems must always be carried out by qualified personnel.

4 Application area

EPL Ga instrument

The VEGABOX 03 is installed in hazardous areas requiring EPL Ga instruments.

EPL Gb instrument

The VEGABOX 03 is installed in hazardous areas requiring EPL Gb instruments.

5 Specific conditions of use ("X" identification)

The following overview is listing all special properties of VEGABOX 03, which make a labelling with the symbol "X" behind the certificate number necessary.

Electrostatic charging (ESD)

You can find the details in chapter " Electrostatic charging (ESD)" of these safety instructions.

Ambient temperature

You can find the details in chapter " Thermal data" of these safety instructions.

Impact and friction sparks

The VEGABOX 03 in light metal versions (e.g. aluminium, titanium, zircon) must be mounted in such a way that sparks from impact and friction between light metals and steel (except stainless steel, if the presence of rust particles can be excluded) cannot occur.

Non-grounded, metallic parts

The resistance between aluminium housing to metal measuring point identification plate is $> 10^9$ Ohm.

The capacitance of the metal measuring point identification plate was measured as follows:



Measurement loop identification label	Capacitance
45 x 23 mm (standard)	21 pF
100 x 30 mm	52 pF
73 x 47 mm	61 pF

Installation

If a cable other than the VEGA connection cable is used as interconnection for signal and power supply circuit and the temperature circuit (PT100-measuring circuit), please make sure that the insulation voltage of at least 500 V AC according to EN 50020 sect. 6.4.12 is maintained and the insulation thickness of the wire insulation is at least 0.25 mm.

6 Important information for mounting and maintenance

General instructions

The following requirements must be fulfilled for mounting, electrical installation, setup and maintenance of the instrument:

- The staff must be qualified according the respective tasks
- The staff must be trained in explosion protection
- The staff must be familiar with the respectively valid regulations, e.g. planning and installation acc. to IEC 60079-14
- Make sure when working on the instrument (mounting, installation, maintenance) that there is no explosive atmosphere present, the supply circuits should be voltage-free, if possible.
- The instrument has to be mounted according to the manufacturer specifications, the Certificate
 of Conformity and the valid regulations and standards
- Modifications on the instrument can influence the explosion protection and hence the safety, therefore repairs are not permitted to be conducted by the end user
- Modifications must only be carried out by employees authorized by VEGA company
- Use only approved spare parts
- Components for installation and connection not included in the approval documents are only permitted if these correspond technically to the latest standard mentioned on the cover sheet. They must be suitable for the application conditions and have a separate certificate. The special conditions of the components must be noted and if necessary, the components must be integrated in the type test. This applies also to the components already mentioned in the technical description.
- Vessel installations and probable flow must be taken into account

Mounting

Keep in mind for instrument mounting

- Mechanical damage on the instrument must be avoided
- Mechanical friction must be avoided
- Close the housing lid (s) up to the stop before starting operating, to ensure the IP protection rating specified on the type label

Maintenance

To ensure the functionality of the device, periodic visual inspection is recommended for:

- Secure mounting
- No mechanical damages or corrosion
- Worn or otherwise damaged cables
- No loose connections of the line connections, equipotential bonding connections
- Correct and clearly marked cable connections



Intrinsic safety "i"

- Valid regulations for connection of intrinsically safe circuits, e.g. proof of intrinsic safety according to IEC/EN 60079-14 must be observed
- The instrument is only suitable for connection to certified, intrinsically safe instruments
- When connecting a circuit with protection level Ex ib, the device, the sensor meas. system of the device must no more be used in hazardous areas of zone 0.
- When connecting an intrinsically safe instruments with classification mark Ex ia to a circuit with
 protection level Ex ib, then the classification mark of the instrument changes to Ex ib. After the
 use as instrument with Ex ib power supply, the instrument must no more be used in circuits with
 protection level Ex ia
- When connecting an intrinsically safe instrument to an non-intrinsically safe circuit, the instrument must be no longer used in intrinsically safe circuits
- With surface temperatures > 70 °C, the cables must be suitable for the higher application conditions

7 Safe operating mode

General operating conditions

- Do not operate the instrument outside the electrical, thermal and mechanical specifications of the manufacturer
- Use the instrument only in media against which the wetted parts are sufficiently resistant
- Note the relation between process temperature on the sensor/antenna and the permissible ambient temperature on the electronics housing. For permissible temperatures, see the respective temperature tables. See chapter " *Thermal data*".
- If necessary, a suitable overvoltage arrester can be connected in front of the VEGABOX 03
- For assessment and reduction of the explosion risk, valid standards such as for example ISO/ EN 1127-1 must be taken into account

8 Potential equalization/Grounding

- Integrate the instruments into the local potential equalisation, e.g. via the internal or external earth terminal
- The potential equalization terminal must be secured against loosening and twisting
- If grounding of the cable screening is necessary, this must be carried out acc. to the valid standards and regulations, e.g. acc. to IEC/EN 60079-14
- The intrinsically safe input and the intrinsically safe output circuits are ground-free. The voltage resistance against ground is min. 500 Veff.
- The supply and signal circuit zwischen dem VEGABOX 03 und dem Sensor should be set up without grounding

9 Electrostatic charging (ESD)

In case of instrument versions with electrostatically chargeable plastic parts, the danger of electrostatic charging and discharging must be taken into account!

The following parts can charge and discharge:

- Lacquered housing version or alternative special lacquering
- Plastic housing, plastic housing parts
- Metal housing with inspection window
- Plastic process fittings
- Plastic-coated process fittings and/or plastic-coated sensors
- Connection cable for separate versions
- Type label



• Isolated metallic labels (measuring point identification plate)

Take note in case of danger of electrostatic charges:

- Avoid friction on the surfaces
- Do not dry clean the surfaces

The instruments must be mounted/installed in such a way that the following can be ruled out:

- electrostatic charges during operation, maintenance and cleaning.
- process-related electrostatic charges, e.g. by measuring media flowing past

The warning label indicates danger:

WARNING - POTENTIAL ELECTROSTATIC CHARGING HAZARD - SEE INSTRUCTIONS

10 Instructions for zone 0 applications

In hazardous areas, the instrument should only be operated under atmospheric conditions:

- Temperature: -20 ... +60 °C.
- Pressure: 80 ... 110 kPa (0.8 ... 1.1 bar)
- Air with normal oxygen content, normally 21 %

If no explosive mixtures or additional application conditions are certified resp. supplementary measures such as e.g. according to EN 1127-1 taken, then the instruments can be also operated according to the manufacturer specification outside atmospheric conditions.

If there is a risk of dangerous potential differences inside zone 0, then suitable measures for circuits in zone 0 must be taken, e.g. according to the requirements of IEC/EN 60079-14.

11 Electrical data

Supply and signal circuit:	
Terminals 1, 2	In type of protection intrinsic safety Ex ia IIC/IIB
	For connection to a certified, intrinsically safe circuit.
	Maximum values:
	● Ui = 30 V ● li = 150 mA ● Pi = 1000 mW ● Ci = 0 ● Li = 0
	When using the supplied connection cable, the following cable in- ductances L' and cable capacitances C' have to be taken into account:
	 Li = 0.6 μH/m C_{1 wire/wire} = 133 pF/m C_{1 wire/screen} = 215 pF/m



Temperature measuring circuit:	
Terminals 3, 4, 5, 6	In type of protection intrinsic safety Ex ia IIC/IIB
	For connection to a certified, intrinsically safe circuit.
	Maximum values:
	• $Ui = 30 V$ • $Ii = 100 mA$ • $Pi = 500 mW$ • $Ci = 0$ • $Li = 0$
	When using the supplied connection cable, the following cable inductances L_i^{\prime} and cable capacitances C_i^{\prime} have to be taken into account:
	 Li = 0.6 μH/m C_{1 wire/wire} = 188 pF/m C_{1 wire/screen} = 555 pF/m

The intrinsically safe circuits are electrically isolated from each other and from parts which can be grounded.

12 Thermal data

In version VEGABOX 03 with terminal blocks

Permissible ambient temperatures depending on temperature class

EPL Ga instrument

Temperature cl	ass	T6 T1
Permissible amb	ient temperature	-20 +60 °C

The connection housing must only be operated in a hazardous area requiring EPL Ga instruments if there are atmospheric conditions (pressure of 0.8 bar to 1.1 bar). If there is no explosive atmosphere, then the permissible operating temperatures and pressures must be taken from the manufacturer specifications.

EPL Gb instrument

Temperature class	Т6 Т1
Permissible ambient temperature	-50 +80 °C

The permissible operating temperatures without explosion-endangered atmosphere are mentioned in the respective manufacturer instructions, e.g. operating instructions manuals.







Printing date:



All statements concerning scope of delivery, application, practical use and operating conditions of the sensors and processing systems correspond to the information available at the time of printing. Subject to change without prior notice

© VEGA Grieshaber KG, Schiltach/Germany 2021

VEGA Grieshaber KG Am Hohenstein 113 77761 Schiltach Germany

Phone +49 7836 50-0 E-mail: info.de@vega.com www.vega.com