

#### Reliable

Certified materials according to FDA and EC 1935/2004 regulations

# **Cost effective**

plics® concept: short delivery time and standardized operation

#### **User friendly**

One measuring system, three measured values: level, pressure, temperature



# Beer tank

## Level measurement in the beer tank

During the brewing process, a thick layer of foam is always present in the beer tanks. Therefore usually two pressure transmitters are used for level measurement. One transmitter measures the head pressure, while the other at the bottom measures the total pressure. The level is precisely calculated from the differential between these two pressures. Hygienic process fittings are absolutely necessary for these sensors.

## More details



### **VEGABAR 82**

Electronic differential pressure measurement for determining the level in the beer tank

- Temperature resistant, CIP capable, linear, overload proof and hysteresisfree
- Reliable level measurement, independent of foaming
- CERTEC® measuring cell of sapphire-ceramic® with gap-free surface meets the highest hygienic standards

#### **Show Product**



	VEGABAR 82
	Show Product
Measuring ra -	nge - Distance
Measuring ra	nge - Pressure
-1 100 bar	
Process temp	perature
-40 150 °C	-
Process pres	SUIP
-1 100 bar	0410
Accuracy 0.05 %	
Materials, we	tted parts
PVDF 316L	
Alloy C22 (2.4)	302)
PP	552)
1.4057	
1.4410	
Alloy C276 (2.4	4819)
Duplex (1.4462	2)
Titanium Grade	e 2 (3.7035)
Threaded cor	nection
≥ G½, ≥ ½ NP	Г
Flange conne	ection
≥ DN15, ≥ ½"	
Hygenic fittin	gs
Clamp ≥ 1" - D	IN32676, ISO2852
Slotted nut ≥ D	N25 - DIN 11851
hygienic fitting	with tension flange DN32
	F40 with compression nut
DRD connection	
SMS 1145 DN SMS DN38	51
SMS DN38 Swagelok VCF	2 screwing
Varivent G125	( Solowing
Varivent N50-4	0
for NEUMO Bi	oControl D50 PN16 / 316L
Seal material	
EPDM	
FKM	

FKM FFKM

